
PRODUCT ALERT

Clinical Center Standardization Committee



MAY 2003 (1)

ROUTINE SPECIMEN BAG

SITUATION

Two recent ORS describe defective transmittal sheet pockets in the Routine Specimen Bag (Item #02394).

The product in question was delivered to MMD Nurse Consult Service.



MMD RESPONSE

- The Routine Specimen Bags were examined. Some were defective.
- All other unreleased product was inspected. Some defective product was found.
- We are shipping all unreleased product to the manufacturer for replacement.

RECOMMENDED ACTION FOR OUR CUSTOMERS

- OK to use Routine Specimen Bags if the pocket is intact.
 - If there is a problem, file ORS, use another Routine Specimen Bag, and send defective product labeled with ORS Number to MMD Nurse Consult Service, 10/ B1N238.
-