

PRODUCT ALERT

Clinical Center Standardization Committee



Internal Recall 6ML Green Sodium Heparin Vacutainer Tube Lot September 2003 (2)

ISSUE:

A foreign object was found in one 6ML green sodium heparin vacutainer tube, item ID #00661, lot #2170889.

MANAGEMENT:

1. Inspect each 6ML green sodium heparin vacutainer tube prior to use.
2. OK to use if no foreign object identified.
3. If foreign object identified:
 - a. do not use the tube
 - b. file ORS to include lot number
 - c. return product to MMD Nurse Consult Service, room B1N238.
4. MMD is locating substitute product with a different lot number. MMD will switch out all product with lot #2170889 as soon as substitute product is available.

"Thank you for your support and cooperation."
